

# Total Base Number (TBN) Forward Titration (ASTM D2896-11)

DOC316.52.93101

Potentiometric titration

Application: Petrochemical oils and lubricants

## 1. Introduction

Oils and lubricants have a base reserve designed to neutralize the acids produced after the combustion process to avoid corrosion of engine components. A low Total Base Number (TBN) indicates that the oil has to be changed.

## 2. Principle

TBN is determined by an acid-base titration using  $\text{HClO}_4$  as the titrant.

TBN is expressed in mg of KOH per g of sample.

Since samples are non-aqueous, they are diluted in a mix of chlorobenzene and acetic acid. The solvent for  $\text{HClO}_4$  is acetic acid.

## 3. Electrodes and reagents

**Electrodes:** Glass pH electrode, PHG311-9 + CL114 cable  
Red Rod reference electrode, double junction, REF251. The outer compartment of the REF251 reference electrode has to be filled with a saturated sodium perchlorate solution  
Legacy adapter: the working electrode PHG311 is plugged on the BNC socket using the CL114 cable, and the reference electrode is plugged on the banana socket. Configure the legacy adapter with the parameters that follow:

|                     |                                                      |
|---------------------|------------------------------------------------------|
| Parameter           | Select pH                                            |
| Sensor Name         | Enter the name of the sensor (e.g., PHG311)          |
| Default Temperature | Enter the ambient temperature (e.g., 25 °C or 77 °F) |
| pHiso               | Enter the iso pH value: 6.65                         |

**Titant:**  $\text{HClO}_4$  0.1 eq/L: It is recommended to buy a commercial 0.1 N perchloric acid solution. Otherwise, mix 8.5 mL of 70 to 72% perchloric acid with 500 mL of glacial acetic acid and 30 mL of acetic anhydride. Dilute to 1 L with glacial acetic acid and allow the solution to stand for 24h

**Titration Solvent:** Chlorobenzene + acetic acid; Add one volume of glacial acetic acid to two volumes of chlorobenzene

**Filling solution for the reference electrode:** Saturated solution of  $\text{NaClO}_4$  in glacial acetic acid

**Standard for titrant calibration:** Potassium hydrogen phthalate

Deionized water

## 4. Ranges and settings

### 4.1. Default parameters

The working procedure is described using the following parameters:

- $m \text{ sample (g)} = \frac{10}{\text{expected BN (mg/g)}}$  weighed with the following precision:

| Sample weight (g) | Weighing precision (g) |
|-------------------|------------------------|
| 5 – 10            | 0.02                   |
| 1 – 5             | 0.005                  |
| 0.25 – 1.0        | 0.001                  |
| 0.1 – 0.25        | 0.0005                 |

- Burette volume = 10 mL

#### 4.2. Working range

The sample weights and precisions given in section 4.1 allow the determination of base numbers between 1 and 100 mg/g.

However ASTM D2896-11 standard specifies that this test method can be used to determine base numbers higher than 300 mg/g. In this case, weigh 0.1 g of sample for expected base numbers up to 450 mg/g. For even more concentrated samples, weigh a smaller amount of sample.

#### 4.3. Settings

| Name                             | Default parameter        | Unit    |
|----------------------------------|--------------------------|---------|
| <b>Application</b>               |                          |         |
| Application name                 | TBN (HClO <sub>4</sub> ) |         |
| <b>Sample</b>                    |                          |         |
| Name                             | Sample                   |         |
| Amount                           | 2.000                    | [g]     |
| <b>Probe</b>                     |                          |         |
| Recommended probe                | PHG311-9                 |         |
| <b>Titrant</b>                   |                          |         |
| Name                             | HClO <sub>4</sub>        |         |
| Titrant concentration            | 0.1000                   | [eq/L]  |
| Syringe                          | Syringe 1                |         |
| <b>Rinsing step 1 (solvent)</b>  |                          |         |
| Active                           | Yes                      |         |
| Time                             | 30                       | [s]     |
| Stirring speed                   | 25                       | [%]     |
| <b>Rinsing step 2 (DI water)</b> |                          |         |
| Active                           | Yes                      |         |
| Time                             | 30                       | [s]     |
| Stirring speed                   | 25                       | [%]     |
| <b>Rinsing step 3 (solvent)</b>  |                          |         |
| Active                           | Yes                      |         |
| Time                             | 30                       | [s]     |
| Stirring speed                   | 25                       | [%]     |
| <b>IP titration</b>              |                          |         |
| Stirring speed                   | 25                       | [%]     |
| Measured parameter               |                          | [mV]    |
| Predose                          | 0                        | [mL]    |
| Max volume stop point            | 10                       | [mL]    |
| Stop on last EQP                 | Yes                      |         |
| Delay                            | 15                       | [s]     |
| Min increment size               | 0.08                     | [mL]    |
| Max increment size               | 0.25                     | [mL]    |
| Result 1 name                    | TBN (mgKOH/g)            |         |
| R1 resolution                    | 2 decimals               |         |
| R1 min                           | 0.01                     | [mg/g]  |
| R1 max                           | 300                      | [mg/g]  |
| R1 QC min                        | 0.01                     | [mg/g]  |
| R1 QC max                        | 300                      | [mg/g]  |
| R1EQP index                      | 1                        |         |
| R1 molar weight                  | 56.11                    | [g/mol] |

#### 4.4. Modification of the settings

The parameters are defined to have the best compromise between accuracy and titration time.

For higher concentration with a high titrant volume, titration time can be reduced with an addition of titrant (predose) at the beginning of the titration. Enter the predose volume (in mL) and the stirring time after the addition in the application edit window.

### 5. Titration procedure

---

A titration is divided into several parts:

- A rinsing procedure for the electrodes, comprising three steps: 30 seconds in solvent (chlorobenzene + acetic acid), 30 seconds in deionized water and another 30 seconds in solvent.
- A blank determination, to determine the equivalent volume induced by the solvent.
- A second rinsing procedure.
- A titration of Total Base Number.

#### 5.1. Sample analysis

Launch the application **TBN (HClO<sub>4</sub>)**.

On the first screen, in **Sample type** choose **Define blank** and press **Start**.

Follow the rinsing indications on the screen. Then, place an empty beaker with a stir bar under the probe holder, add 60 mL of solvent with a graduated cylinder and make sure that both electrodes are immersed. At the end of the titration, the equivalent volume corresponding to the blank is displayed and automatically recorded. Press **Next** and chose **New sample**.

**Note:** For each set of samples make a blank on 60 mL of titration solvent.

Weigh the sample in the titration beaker according to the recommendations given in section 4.1.

In **Sample type** choose **Sample with blank** (or **Sample** if blank was 0 mL) and press **Start**.

Follow the rinsing indications on the screen. Then, place the beaker containing the sample and a stir bar under the probe holder, press **OK**. Add 60 mL of solvent with a graduated cylinder, making sure that both electrodes are immersed and press **OK**. Titration will start after a 15-second delay. At the end of the titration, TBN is displayed in mg of KOH per g of sample.

By pressing **Next** it is possible to:

- Replicate the sample. This is used to study the repeatability by analyzing several samples successively. At the end of each titration, a window displays the average value, the standard deviation (SD in mg/g) and the relative standard deviation (RSD in %).
- Analyze a new sample. Another titration can be started but no Standard Deviation and RSD value will be made.

If no inflection is detected then it is possible to use back titration (see the corresponding Application file and Working procedure).

### 6. Results

---

#### 6.1. Result calculation

The calculation used is:

$$\begin{aligned} \text{TBN (mgKOH/g)} &= \frac{C_{\text{titrant}} (\text{eq/L}) \times (V_{\text{titrant}} - V_{\text{blank}}) (\text{mL})}{n_{\text{e- titrant}} \times m_{\text{sample}} (\text{g})} \times M_{\text{KOH}} (\text{g/mol}) \\ &= \frac{0.1 (\text{eq/L}) \times (V_{\text{titrant}} - V_{\text{blank}}) (\text{mL})}{1 \times m_{\text{sample}} (\text{g})} \times 56.11 (\text{g/mol}) \end{aligned}$$

## 6.2. Experimental results

These results are indicative and have been obtained for a given sample for six successive determinations. Results are given for the titration of 3 standards and a sample of used motor oil.

| Standard 1.01 mg/g              |      |
|---------------------------------|------|
| Mean TBN (mg/g)                 | 0.98 |
| Standard deviation (mg/g)       | 0.01 |
| Relative standard deviation (%) | 0.4  |

| Standard 6.06 mg/g              |      |
|---------------------------------|------|
| Mean TBN (mg/g)                 | 6.07 |
| Standard deviation (mg/g)       | 0.05 |
| Relative standard deviation (%) | 0.7  |


| Standard 40.40 mg/g             |       |
|---------------------------------|-------|
| Mean TBN (mg/g)                 | 39.78 |
| Standard deviation (mg/g)       | 0.14  |
| Relative standard deviation (%) | 0.4   |

| Used motor oil                  |      |
|---------------------------------|------|
| Mean TBN (mg/g)                 | 9.73 |
| Standard deviation (mg/g)       | 0.04 |
| Relative standard deviation (%) | 0.5  |

The titration typically lasts around 2 min.

## 6.3. Example of a titration curve

This curve has been obtained during the analysis of one of the samples.



## 7. Recommendations

Refill the reference electrode regularly to maintain the level of both internal solutions (inner and outer) at around 1 cm (0.4 inches) below the refill hole.

Rinsing steps can be skipped (right button) or disabled (in Edit mode), although it is highly recommended to consistently rinse the electrodes between each measurement to maintain good accuracy.

## 8. Bibliography

- *Standard ASTM D2896-11*

## 9. Appendix: Titrant calibration

### 9.1. Electrode testing and storage

It is recommended to check the electrode's behavior when first put into use, or when new electrodes are installed, and retest at intervals thereafter. Dip the electrodes into a well stirred mixture of 60 mL of glacial acetic acid plus 0.1 g of potassium hydrogen phthalate and record the reading (**Maintenance > Live Measure**).

Rinse the electrodes with chlorobenzene and immerse in a 50 mL of glacial acetic acid plus 0.75 mL of 0.1 N HClO<sub>4</sub> solution. The difference between the readings has to be at least 0.3 V.

When not in use, store both electrodes in deionized water.

### 9.2. Titrant calibration

The perchloric acid solution can be calibrated. Its exact concentration can be determined from an acid-base titration using potassium hydrogen phthalate.

Heat a quantity of potassium hydrogen phthalate in an oven at 120 °C for 2 hours and allow it to cool. Weigh 80 mg of potassium hydrogen phthalate powder and dissolve it in 20 mL of warm glacial acetic acid. Add 40 mL of chlorobenzene and cool. Put in a stir bar, dip the probes and the delivery tip into the solution and launch the titrant calibration sequence. When prompted, type in the exact weighed amount of powder.

At the end of the titrant calibration, titer (eq/L) is displayed and the user can reject, replicate, or save the result. The saved value will be used for calculations.

#### Default settings for titrant calibration

| Name                       | Default parameter            | Unit    |
|----------------------------|------------------------------|---------|
| <b><i>Titrant</i></b>      |                              |         |
| Name                       | HClO <sub>4</sub>            |         |
| Titrant concentration      | 0.1000                       | [eq/L]  |
| Syringe                    | Syringe 1                    |         |
| <b><i>Standard</i></b>     |                              |         |
| Name                       | Potassium hydrogen phthalate |         |
| Amount                     | 80                           | [mg]    |
| Amount min                 | 50                           | [mg]    |
| Amount max                 | 100                          | [mg]    |
| Molar weight               | 204.22                       | [g/mol] |
| <b><i>IP titration</i></b> |                              |         |
| Stirring speed             | 30                           | [%]     |
| Measured parameter         |                              | [mV]    |
| Predose                    | 1.5                          | [mL]    |
| Max volume stop point      | 10                           | [mL]    |
| Stop on last EQP           | True                         |         |
| Delay                      | 0                            | [s]     |
| Min increment size         | 0.05                         | [mL]    |
| Max increment size         | 0.2                          | [mL]    |
| Result name                | Titer                        |         |
| Result resolution          | 4 decimals                   |         |
| Result min                 | 0.09                         | [eq/L]  |
| Result max                 | 0.11                         | [eq/L]  |